

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 13 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130301 ✓
Name	Kyle Carrington	ID#	5336 Date 7-8-09 ✓
A	Agency	UNFPD	Phone # 474-7721
Instrument Location		612 Yukon Drive Fairbanks, AK 99775 ✓	
Alco S/N		77198 ✓	Target Value .075 ✓ High Pressure 400 PSI
B		Alco Test Values	
		1 st Alco .078 ✓	2 nd Alco .078 ✓
Signature		Kyle Carrington ✓	
		(OVER)	

CDB
7/14/09


(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Jul, 2009.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130301

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

JULY 08, 2009

OPERATOR'S NAME:

CARRINGTON, KYLE

OPERATOR'S NUMBER: 5336

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

I

O.L. #: 0

DEPT/AGENCY: FBK7

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .075

ALCO S/N: 77198

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.75 in
ALCO TARGET .074 09:50
BLANK TEST .000 09:51
INTERNAL STANDARD .078 VERIFIED 09:51
ALCO TU 29.75 in .078 09:51
BLANK TEST .000 09:52
SUBJECT SAMPLE .000 09:53
BLANK TEST .000 09:54
ALCO TU 29.75 in .078 09:54
BLANK TEST .000 09:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

JULY 08, 2009

TIME 09:47

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.75 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abdefghijklmno
pqrstuvwxyz{|}~